
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000223
7:06 28/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:29
STERILIZING 5:08
136.0 C 228kPa 5:08
136.3 C 231kPa 5:38
136.4 C 231kPa 6:08
136.3 C 232kPa 6:38
136.2 C 230kPa 7:08
136.4 C 232kPa 7:38
136.4 C 232kPa 8:08
136.4 C 232kPa 8:38
136.4 C 232kPa 8:38
VENTING 8:38
AIR DRYING 9:38
CYCLE COMPLETE 11:32

Digital Signature #
67F141CFC16CBB04
