
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 001532
14:22 10/25/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:54
PRESSURIZING 6:36
STERILIZING 7:57
136.0 C 220kPa 7:57
136.0 C 219kPa 8:27
136.4 C 224kPa 8:58
136.2 C 221kPa 9:28
136.4 C 223kPa 9:58
136.4 C 223kPa 10:28
136.3 C 222kPa 10:58
136.2 C 221kPa 11:27
VENTING 11:28
AIR DRYING 12:31
CYCLE COMPLETE 42:31

Digital Signature #
DBFEABE19202C5E6
