

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4272

Name: RITTER2  
Sterilizer ID: V2313228  
01/16/2025 05:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 124.5 F 0.0  
2:00 127.4 F 0.1  
4:00 152.3 F 0.4  
6:00 187.0 F 2.1  
8:00 206.3 F 6.5  
10:00 220.5 F 11.4  
12:00 226.4 F 7.0  
14:00 239.3 F 12.3  
16:00 238.6 F 10.4  
18:00 236.6 F 9.2  
20:00 251.3 F 16.2  
22:00 263.3 F 23.2  
24:00 271.8 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.8  
1:00 272.6 F 29.2  
2:00 272.2 F 29.3  
3:00 272.4 F 29.0  
4:00 272.4 F 29.5

Min 271.7 F 28.6  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.5 F 0.0  
5:00 194.9 F 0.0  
10:00 217.5 F 0.0  
15:00 225.8 F 0.0  
20:00 227.5 F 0.0  
25:00 229.9 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 24:05  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:02:39  
CYCLE COMPLETE

01/16/2025 06:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_