
Midmark M11 - v1.0.5
Total Cycles: 983

Name: RITTER2
Sterilizer ID: V2313228
04/13/2022 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.3 F 0.0
2:00 208.2 F 2.0
4:00 211.1 F 7.6
6:00 226.4 F 13.5
8:00 233.4 F 10.0
10:00 237.8 F 8.6
12:00 241.3 F 10.9
14:00 247.7 F 14.5
16:00 260.5 F 21.6
18:00 271.9 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.4 F 29.4
2:00 271.7 F 29.1
3:00 272.4 F 29.4
4:00 271.5 F 28.9

Min 271.5 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 216.9 F 0.0
10:00 231.1 F 0.0
15:00 229.8 F 0.0
20:00 228.6 F 0.0
25:00 228.8 F 0.0
30:00 225.0 F 0.0

Drying Complete

FILLING: 1:42
HEATING: 18:33
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:26
CYCLE COMPLETE

04/13/2022 12:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____