

-----  
Midmark M11 - v1.0.5  
Total Cycles: 181

Name: RITTER2  
Sterilizer ID: V2313228  
07/08/2021 09:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.9 F 0.0  
2:00 176.8 F 2.7  
4:00 200.7 F 8.1  
6:00 221.5 F 10.1  
8:00 231.3 F 10.0  
10:00 233.3 F 8.2  
12:00 240.6 F 10.9  
14:00 247.6 F 14.4  
16:00 260.5 F 21.7  
18:00 272.2 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.0 F 29.1  
2:00 272.6 F 29.6  
3:00 271.6 F 29.3  
4:00 272.0 F 29.2  
  
Min 271.2 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.8 F 0.0  
5:00 231.1 F 0.0  
10:00 221.1 F 0.0  
15:00 221.1 F 0.0  
20:00 227.0 F 0.0  
25:00 223.1 F 0.0  
30:00 219.4 F 0.0  
  
Drying Complete

FILLING: 2:44  
HEATING: 18:25  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 00:58:10  
CYCLE COMPLETE

07/08/2021 10:54 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02