

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2977

Name: RITTER2  
Sterilizer ID: V2313228  
10/05/2023 10:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.4 F 0.0  
2:00 157.9 F 2.2  
4:00 171.4 F 5.7  
6:00 195.0 F 9.9  
8:00 207.8 F 6.9  
10:00 225.1 F 7.8  
12:00 236.0 F 12.5  
14:00 235.4 F 9.7  
16:00 237.7 F 9.5  
18:00 247.4 F 14.6  
20:00 259.2 F 21.0  
22:00 269.8 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.7 F 29.5  
2:00 271.1 F 29.0  
3:00 272.5 F 29.6  
4:00 271.4 F 29.0

Min 271.1 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.8 F 0.0  
5:00 233.4 F 0.0  
10:00 231.5 F 0.0  
15:00 228.6 F 0.0  
20:00 230.1 F 0.0  
25:00 228.1 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 22:56  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:01:38  
CYCLE COMPLETE

10/05/2023 11:24 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_