

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002639  
9:56 09/11/2021  
RUBBER AND PLASTICS  
121 C for 15 min.  
CYCLE START 0:00  
CONDITIONING 2:46  
PRESSURIZING 4:12  
STERILIZING 4:34  
123.0 C 116kPa 4:34  
123.4 C 118kPa 5:04  
123.4 C 118kPa 5:34  
123.4 C 116kPa 6:04  
123.3 C 118kPa 6:34  
123.4 C 116kPa 7:04  
123.4 C 118kPa 7:34  
123.4 C 118kPa 8:04  
123.4 C 116kPa 8:34  
123.4 C 116kPa 9:04  
123.4 C 117kPa 9:34  
123.1 C 116kPa 10:04  
123.4 C 118kPa 10:34  
123.4 C 118kPa 11:04  
123.4 C 118kPa 11:34  
123.4 C 118kPa 12:04  
123.4 C 117kPa 12:34  
123.4 C 116kPa 13:04  
123.3 C 119kPa 13:34  
123.4 C 118kPa 14:04  
123.4 C 117kPa 14:34  
123.4 C 116kPa 15:04  
123.4 C 116kPa 15:34  
123.4 C 117kPa 16:04  
123.2 C 117kPa 16:34  
123.5 C 118kPa 17:04

123.2 C 116kPa 17:34  
123.2 C 117kPa 18:04  
123.4 C 115kPa 18:34  
123.4 C 118kPa 19:04  
123.4 C 118kPa 19:34  
123.4 C 118kPa 19:34  
VENTING 19:35  
AIR DRYING 20:19  
CYCLE COMPLETE 20:23

-----  
Digital Signature #  
14D2904C5DD2C8F9  
-----