
Midmark M11 - v1.0.5
Total Cycles: 4060

Name: RITTER2
Sterilizer ID: V2313228
10/28/2024 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.1 F 0.0
2:00 118.5 F 0.0
4:00 123.1 F 0.0
6:00 159.1 F 1.0
8:00 180.4 F 3.9
10:00 200.5 F 7.1
12:00 209.5 F 11.2
14:00 221.0 F 5.5
16:00 228.0 F 9.5
18:00 236.2 F 11.0
20:00 239.4 F 11.4
22:00 238.1 F 9.9
24:00 250.7 F 15.9
26:00 261.0 F 21.8
28:00 270.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.6 F 29.6
2:00 271.9 F 29.0
3:00 271.2 F 29.0
4:00 272.2 F 29.5

Min 271.2 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 185.7 F 0.0
10:00 205.7 F 0.0
15:00 203.1 F 0.0
20:00 202.9 F 0.0
25:00 204.1 F 0.0
30:00 207.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 28:53
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:07:33
CYCLE COMPLETE

10/28/2024 12:39 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____