
Midmark M11 - v1.0.0
Total Cycles: 2355

Name: RITTER 1
Sterilizer ID: V2178198
09/21/2022 06:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.5 F 0.0
2:00 185.9 F 4.5
4:00 210.6 F 9.3
6:00 221.3 F 6.7
8:00 231.5 F 9.8
10:00 236.8 F 8.5
12:00 243.5 F 13.0
14:00 242.3 F 11.7
16:00 255.5 F 18.5
18:00 267.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.7 F 29.4
2:00 272.2 F 28.8
3:00 271.6 F 28.7
4:00 272.7 F 29.5

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 231.6 F 0.0
10:00 230.5 F 0.0
15:00 229.6 F 0.0
20:00 228.3 F 0.0
25:00 232.5 F 0.0
30:00 227.1 F 0.0

Drying Complete

FILLING: 2:01
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

09/21/2022 07:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____