
Midmark M11 - v1.0.0
Total Cycles: 3295

Name: RITTER 1
Sterilizer ID: V2178198
05/29/2023 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.9 F 0.0
2:00 186.9 F 0.0
4:00 199.5 F 4.6
6:00 217.6 F 9.7
8:00 229.1 F 7.8
10:00 236.8 F 11.3
12:00 236.5 F 9.0
14:00 239.7 F 9.5
16:00 249.0 F 14.7
18:00 261.5 F 21.7
20:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 29.0
2:00 271.9 F 28.4
3:00 271.7 F 28.5
4:00 272.6 F 29.0

Min 271.4 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 227.2 F 0.0
10:00 236.3 F 0.0
15:00 230.2 F 0.0
20:00 227.3 F 0.0
25:00 225.6 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 20:26
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

05/29/2023 11:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____