
Midmark M11 - v1.0.0
Total Cycles: 826

Name: RITTER 1
Sterilizer ID: V2178198
05/19/2021 04:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.4 F 0.0
2:00 193.5 F 3.4
4:00 213.4 F 8.2
6:00 227.3 F 13.4
8:00 231.9 F 9.0
10:00 238.7 F 9.9
12:00 245.0 F 12.9
14:00 244.0 F 12.0
16:00 257.0 F 18.8
18:00 268.4 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.3 F 28.4
2:00 272.2 F 28.6
3:00 272.0 F 28.1
4:00 272.4 F 28.8

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 232.9 F 0.0
10:00 230.1 F 0.0
15:00 222.4 F 0.0
20:00 221.2 F 0.0
25:00 221.5 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 2:40
HEATING: 19:10
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

05/19/2021 05:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____