
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003782
14:34 07/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:17
PRESSURIZING 1:21
STERILIZING 2:27
136.0 C 220kPa 2:27
136.4 C 223kPa 2:57
136.3 C 223kPa 3:27
136.3 C 224kPa 3:57
136.2 C 225kPa 4:27
136.1 C 223kPa 4:57
136.4 C 224kPa 5:27
136.2 C 223kPa 5:57
136.2 C 223kPa 5:57
VENTING 5:57
AIR DRYING 6:48
CYCLE COMPLETE 6:51

Digital Signature #
5B6D9D280CF980A0
