
Midmark M11 - v1.0.5
Total Cycles: 2315

Name: RITTER2
Sterilizer ID: V2313228
05/01/2023 07:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.8 F 0.0
2:00 76.8 F 0.0
4:00 112.9 F 0.0
6:00 113.9 F 0.0
8:00 139.3 F 1.5
10:00 169.0 F 4.8
12:00 187.9 F 8.5
14:00 208.7 F 12.2
16:00 218.3 F 5.9
18:00 229.3 F 9.9
20:00 233.4 F 10.1
22:00 238.9 F 11.5
24:00 236.7 F 9.8
26:00 249.8 F 15.9
28:00 260.4 F 22.1
30:00 270.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 271.9 F 30.0
2:00 272.0 F 29.6
3:00 271.3 F 29.6
4:00 272.3 F 30.0

Min 271.1 F 29.1
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 198.6 F 0.0
10:00 215.8 F 0.0
15:00 218.5 F 0.0
20:00 225.9 F 0.0
25:00 228.4 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:43
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:46
CYCLE COMPLETE

05/01/2023 08:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____