
Midmark M11 - v1.0.5
Total Cycles: 3424

Name: RITTER2
Sterilizer ID: V2313228
02/29/2024 03:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.9 F 0.0
2:00 147.3 F 1.8
4:00 162.9 F 5.0
6:00 186.1 F 9.6
8:00 211.7 F 6.5
10:00 224.2 F 8.9
12:00 235.6 F 10.4
14:00 239.7 F 11.8
16:00 239.8 F 10.5
18:00 253.2 F 17.3
20:00 264.8 F 24.2
22:00 271.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 271.7 F 29.4
2:00 271.7 F 29.5
3:00 271.6 F 29.3
4:00 271.8 F 29.2

Min 271.5 F 29.2
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.8 F 0.0
5:00 246.4 F 0.0
10:00 190.8 F 0.0
15:00 187.9 F 0.0
20:00 184.6 F 0.0
25:00 182.3 F 0.0
30:00 180.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:28
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:00:49
CYCLE COMPLETE

02/29/2024 04:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____