
Midmark M11 - v1.0.5
Total Cycles: 1486

Name: RITTER2
Sterilizer ID: V2313228
09/13/2022 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.1 F 0.0
2:00 150.9 F 2.2
4:00 162.2 F 6.0
6:00 189.2 F 11.2
8:00 215.1 F 6.1
10:00 230.4 F 11.4
12:00 231.5 F 8.6
14:00 239.1 F 10.5
16:00 245.9 F 14.1
18:00 258.5 F 21.0
20:00 269.8 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.5 F 29.7
2:00 272.8 F 30.2
3:00 271.5 F 29.4
4:00 272.4 F 30.0

Min 271.5 F 29.3
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 214.4 F 0.0
10:00 211.4 F 0.0
15:00 228.9 F 0.0
20:00 229.4 F 0.0
25:00 226.0 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:58
CYCLE COMPLETE

09/13/2022 02:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____