

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2413

Name: RITTER2  
Sterilizer ID: V2313228  
05/25/2023 08:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.7 F 0.0  
2:00 190.2 F 2.7  
4:00 192.0 F 7.5  
6:00 214.2 F 13.2  
8:00 227.8 F 9.2  
10:00 232.9 F 9.0  
12:00 243.8 F 13.2  
14:00 243.9 F 12.4  
16:00 257.1 F 19.4  
18:00 268.8 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 272.5 F 29.5  
2:00 272.3 F 29.9  
3:00 272.2 F 29.2  
4:00 272.9 F 29.8  
  
Min 271.5 F 29.1  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.1 F 0.0  
5:00 224.5 F 0.0  
10:00 222.5 F 0.0  
15:00 238.6 F 0.0  
20:00 228.2 F 0.0  
25:00 227.3 F 0.0  
30:00 228.4 F 0.0  
  
Drying Complete

FILLING: 1:53  
HEATING: 19:51  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:55  
CYCLE COMPLETE

05/25/2023 09:49 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_