
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
2.6uS / 1.6ppm
CYCLE NUMBER 000779
11:47 17/02/2022
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CYCLE ABORTED 0:03
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:03

Digital Signature #
C0565D1AEC9A941A
