
Midmark M11 - v1.0.5
Total Cycles: 1166

Name: RITTER2
Sterilizer ID: V2313228
06/07/2022 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.8 F 0.0
2:00 184.2 F 3.5
4:00 195.6 F 8.5
6:00 216.9 F 9.1
8:00 227.7 F 9.8
10:00 233.7 F 8.7
12:00 243.1 F 12.7
14:00 243.5 F 13.0
16:00 256.7 F 19.9
18:00 268.0 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.5 F 30.0
2:00 271.6 F 29.5
3:00 272.6 F 30.0
4:00 271.8 F 29.4

Min 271.6 F 29.3
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 208.3 F 0.0
10:00 217.4 F 0.0
15:00 221.8 F 0.0
20:00 222.5 F 0.0
25:00 223.1 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 2:27
HEATING: 19:16
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

06/07/2022 11:07 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____