

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5111

Name: RITTER 1  
Sterilizer ID: V2178198  
11/19/2024 04:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.8 F 0.0  
2:00 141.9 F 0.0  
4:00 159.4 F 0.4  
6:00 187.5 F 1.1  
8:00 204.7 F 4.9  
10:00 218.0 F 9.1  
12:00 229.8 F 13.5  
14:00 233.3 F 9.2  
16:00 238.9 F 10.5  
18:00 243.6 F 12.4  
20:00 241.8 F 11.2  
22:00 254.3 F 17.5  
24:00 265.2 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.4 F 29.0  
2:00 272.1 F 28.6  
3:00 271.4 F 28.1  
4:00 271.6 F 28.5

Min 271.3 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.8 F 0.0  
5:00 199.3 F 0.0  
10:00 202.8 F 0.0  
15:00 207.7 F 0.0  
20:00 220.3 F 0.0  
25:00 227.4 F 0.0  
30:00 228.4 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 25:58  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:04:34  
CYCLE COMPLETE

11/19/2024 05:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_