
Midmark M11 - v1.0.0
Total Cycles: 3803

Name: RITTER 1
Sterilizer ID: V2178198
09/21/2023 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 73.5 F 0.0
4:00 86.8 F 0.0
6:00 103.3 F 0.0
8:00 142.7 F 0.4
10:00 162.1 F 2.1
12:00 181.9 F 4.0
14:00 194.1 F 6.3
16:00 207.1 F 9.2
18:00 220.5 F 12.6
20:00 224.6 F 6.5
22:00 234.0 F 10.6
24:00 234.9 F 8.4
26:00 243.5 F 12.3
28:00 241.1 F 10.7
30:00 253.0 F 16.6
32:00 263.4 F 22.7
34:00 272.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.5 F 28.6
2:00 271.4 F 28.0
3:00 272.5 F 28.7
4:00 271.6 F 28.1

Min 271.4 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.8 F 0.0
5:00 199.6 F 0.0
10:00 198.0 F 0.0
15:00 197.2 F 0.0
20:00 188.6 F 0.0
25:00 189.3 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 34:15
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:12:39
CYCLE COMPLETE

09/21/2023 09:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____