

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4482

Name: RITTER2  
Sterilizer ID: V2313228  
04/03/2025 10:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.8 F 0.0  
2:00 162.6 F 1.6  
4:00 172.4 F 4.7  
6:00 185.6 F 9.0  
8:00 208.7 F 9.7  
10:00 222.8 F 7.0  
12:00 233.9 F 11.6  
14:00 234.0 F 8.5  
16:00 241.5 F 11.4  
18:00 243.7 F 13.1  
20:00 255.7 F 19.4  
22:00 266.7 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.5  
1:00 271.8 F 30.2  
2:00 272.4 F 29.8  
3:00 271.3 F 29.8  
4:00 272.5 F 30.2

Min 271.2 F 29.4  
Max 272.7 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.5 F 0.0  
5:00 195.4 F 0.0  
10:00 200.7 F 0.0  
15:00 201.2 F 0.0  
20:00 199.9 F 0.0  
25:00 199.8 F 0.0  
30:00 199.0 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 23:28  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:02:15  
CYCLE COMPLETE

04/03/2025 11:32 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_