

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000448  
10:41 15/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:44  
PRESSURIZING 3:45  
STERILIZING 4:52  
136.4 C 223kPa 4:52  
136.1 C 222kPa 5:22  
136.2 C 223kPa 5:52  
136.3 C 224kPa 6:22  
136.1 C 223kPa 6:52  
136.4 C 224kPa 7:22  
136.4 C 225kPa 7:52  
136.4 C 224kPa 8:22  
136.4 C 224kPa 8:22  
VENTING 8:22  
AIR DRYING 9:19  
CYCLE COMPLETE 9:25

-----  
Digital Signature #  
B788458EF1A131D7  
-----