
Midmark M11 - v1.0.5
Total Cycles: 3577

Name: RITTER2
Sterilizer ID: V2313228
05/01/2024 02:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.9 F 0.0
2:00 149.0 F 2.7
4:00 171.3 F 6.4
6:00 194.3 F 10.6
8:00 211.5 F 5.1
10:00 222.7 F 8.1
12:00 233.9 F 12.7
14:00 235.9 F 10.0
16:00 237.3 F 9.4
18:00 248.4 F 15.0
20:00 259.7 F 21.4
22:00 270.2 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.9 F 30.0
2:00 272.4 F 29.5
3:00 271.5 F 29.1
4:00 271.8 F 29.7

Min 271.4 F 29.1
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.4 F 0.0
5:00 198.5 F 0.0
10:00 193.2 F 0.0
15:00 187.8 F 0.0
20:00 184.2 F 0.0
25:00 181.4 F 0.0
30:00 178.5 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:58
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

05/01/2024 03:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____