

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002432  
8:50 22/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:47  
PRESSURIZING 2:49  
STERILIZING 3:53  
136.4 C 224kPa 3:53  
136.4 C 224kPa 4:23  
136.1 C 220kPa 4:53  
136.2 C 223kPa 5:23  
136.4 C 223kPa 5:53  
136.4 C 224kPa 6:23  
136.2 C 224kPa 6:53  
136.2 C 223kPa 7:23  
136.2 C 223kPa 7:23  
VENTING 7:23  
AIR DRYING 8:14  
CYCLE COMPLETE 31:45  
-----

Digital Signature #  
4E7A00EA392A7FBC  
-----