

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2892

Name: RITTER2  
Sterilizer ID: V2313228  
09/15/2023 02:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.6 F 0.0  
2:00 170.4 F 2.9  
4:00 188.4 F 7.0  
6:00 207.1 F 11.4  
8:00 224.6 F 5.9  
10:00 228.8 F 10.8  
12:00 231.3 F 7.6  
14:00 241.9 F 12.4  
16:00 244.7 F 12.8  
18:00 256.5 F 19.1  
20:00 267.1 F 25.7  
22:00 272.5 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.4 F 29.1  
1:00 272.6 F 29.4  
2:00 272.3 F 29.7  
3:00 271.8 F 29.5  
4:00 271.9 F 29.1

Min 271.5 F 29.0  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
5:00 223.1 F 0.0  
10:00 195.7 F 0.0  
15:00 190.3 F 0.0  
20:00 183.7 F 0.0  
25:00 195.7 F 0.0  
30:00 198.4 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 22:01  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:01:26  
CYCLE COMPLETE

09/15/2023 03:53 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_