
Midmark M11 - v1.0.0
Total Cycles: 1108

Name: RITTER 1
Sterilizer ID: V2178198
08/23/2021 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.4 F 0.0
2:00 117.8 F 0.0
4:00 141.0 F 0.2
6:00 174.3 F 2.9
8:00 196.0 F 6.6
10:00 213.3 F 11.3
12:00 226.0 F 5.7
14:00 235.3 F 10.9
16:00 233.3 F 7.7
18:00 243.3 F 11.8
20:00 245.5 F 12.8
22:00 258.2 F 19.6
24:00 269.5 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.7 F 28.8
2:00 271.5 F 28.0
3:00 272.4 F 28.9
4:00 272.4 F 28.6

Min 271.2 F 27.9
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 165.7 F 0.0
5:00 192.2 F 0.0
10:00 190.2 F 0.0
15:00 188.5 F 0.0
20:00 184.8 F 0.0
25:00 182.2 F 0.0
30:00 181.5 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 25:01
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:03:20
CYCLE COMPLETE

08/23/2021 09:51 AM
ACCEPTED: X
REJECTED:
Operator: U02