
Midmark M11 - v1.0.0
Total Cycles: 4841

Name: RITTER 1
Sterilizer ID: V2178198
08/22/2024 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 74.0 F 0.0
4:00 89.4 F 0.0
6:00 121.8 F 0.3
8:00 153.0 F 0.7
10:00 183.3 F 1.4
12:00 202.2 F 4.7
14:00 215.3 F 8.0
16:00 226.7 F 11.5
18:00 225.6 F 5.6
20:00 237.8 F 10.5
22:00 234.0 F 7.6
24:00 245.2 F 12.9
26:00 242.2 F 11.1
28:00 254.8 F 17.5
30:00 265.7 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.6 F 28.6
2:00 271.6 F 28.1
3:00 272.4 F 28.6
4:00 271.5 F 28.0

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 215.1 F 0.0
10:00 234.2 F 0.0
15:00 201.9 F 0.0
20:00 196.9 F 0.0
25:00 195.8 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 0:41
HEATING: 31:47
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:09:48
CYCLE COMPLETE

08/22/2024 09:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____