
Midmark M11 - v1.0.0
Total Cycles: 3785

Name: RITTER 1
Sterilizer ID: V2178198
09/18/2023 08:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 88.1 F 0.0
4:00 74.5 F 0.0
6:00 103.4 F 0.2
8:00 139.9 F 1.8
10:00 168.1 F 3.5
12:00 187.8 F 6.4
14:00 203.6 F 9.8
16:00 222.1 F 9.7
18:00 226.0 F 6.8
20:00 235.2 F 10.9
22:00 231.8 F 7.3
24:00 244.1 F 12.6
26:00 241.1 F 10.8
28:00 253.7 F 17.1
30:00 264.7 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.2 F 28.5
2:00 271.6 F 28.5
3:00 272.0 F 28.4
4:00 271.7 F 28.5

Min 271.2 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 221.3 F 0.0
10:00 187.6 F 0.0
15:00 170.3 F 0.0
20:00 172.4 F 0.0
25:00 174.0 F 0.0
30:00 192.6 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:57
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:06
CYCLE COMPLETE

09/18/2023 09:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____