
Midmark M11 - v1.0.0
Total Cycles: 840

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2021 06:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.7 F 0.0
2:00 77.0 F 0.0
4:00 88.3 F 0.0
6:00 113.5 F 0.0
8:00 147.2 F 0.1
10:00 171.1 F 0.2
12:00 191.7 F 0.3
14:00 207.9 F 3.2
16:00 215.3 F 5.9
18:00 225.0 F 8.9
20:00 234.3 F 12.4
22:00 232.6 F 8.1
24:00 243.5 F 12.9
26:00 241.2 F 10.9
28:00 238.2 F 9.3
30:00 251.1 F 15.5
32:00 261.7 F 21.5
34:00 271.5 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.6 F 28.7
2:00 272.2 F 28.2
3:00 271.5 F 27.8
4:00 272.5 F 28.7

Min 271.5 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 193.4 F 0.0
10:00 190.3 F 0.0
15:00 186.3 F 0.0
20:00 187.1 F 0.0
25:00 186.9 F 0.0
30:00 193.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 34:43
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:13:08
CYCLE COMPLETE

05/27/2021 07:53 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____