

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1255

Name: RITTER 1  
Sterilizer ID: V2178198  
10/12/2021 11:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.4 F 0.0  
2:00 183.2 F 4.0  
4:00 199.4 F 9.3  
6:00 230.3 F 7.7  
8:00 232.4 F 11.6  
10:00 235.7 F 9.2  
12:00 238.5 F 8.8  
14:00 248.6 F 14.6  
16:00 261.2 F 21.6  
18:00 272.4 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.9  
1:00 271.8 F 29.0  
2:00 271.8 F 29.1  
3:00 271.7 F 29.0  
4:00 271.6 F 28.9  
  
Min 271.6 F 28.8  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.4 F 0.0  
5:00 236.2 F 0.0  
10:00 226.0 F 0.0  
15:00 226.9 F 0.0  
20:00 226.2 F 0.0  
25:00 227.1 F 0.0  
30:00 218.1 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 19:02  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 00:57:55  
CYCLE COMPLETE

10/12/2021 12:03 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01