

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1029

Name: RITTER2  
Sterilizer ID: V2313228  
04/25/2022 03:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.4 F 0.0  
2:00 149.1 F 2.4  
4:00 164.4 F 6.0  
6:00 191.6 F 10.8  
8:00 213.9 F 4.9  
10:00 226.2 F 10.5  
12:00 230.4 F 7.4  
14:00 241.7 F 13.5  
16:00 242.6 F 12.4  
18:00 255.2 F 19.0  
20:00 266.7 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.5  
1:00 272.3 F 29.6  
2:00 272.4 F 30.1  
3:00 271.5 F 29.5  
4:00 272.4 F 30.0  
  
Min 271.5 F 29.4  
Max 272.5 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.5 F 0.0  
5:00 201.6 F 0.0  
10:00 217.6 F 0.0  
15:00 219.2 F 0.0  
20:00 220.2 F 0.0  
25:00 219.5 F 0.0  
30:00 220.6 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 21:30  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:00  
CYCLE COMPLETE

04/25/2022 04:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_