
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000883
16:25 26/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:29
STERILIZING 4:45
136.0 C 228kPa 4:45
136.2 C 231kPa 5:15
136.1 C 230kPa 5:45
136.4 C 231kPa 6:15
136.3 C 232kPa 6:45
136.3 C 231kPa 7:15
136.4 C 233kPa 7:45
136.2 C 231kPa 8:15
136.2 C 231kPa 8:15
VENTING 8:15
AIR DRYING 9:07
CYCLE COMPLETE 9:10

Digital Signature #
C5DEF5A4996AAF67
