
Midmark M11 - v1.0.0
Total Cycles: 3128

Name: RITTER 1
Sterilizer ID: V2178198
04/19/2023 10:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.3 F 0.0
2:00 199.2 F 0.0
4:00 217.8 F 5.7
6:00 228.1 F 11.2
8:00 230.2 F 7.6
10:00 243.8 F 12.9
12:00 243.5 F 12.2
14:00 242.9 F 11.5
16:00 256.7 F 18.7
18:00 268.8 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.4 F 28.5
2:00 272.0 F 28.5
3:00 272.2 F 28.4
4:00 271.9 F 28.5

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 228.9 F 0.0
10:00 234.1 F 0.0
15:00 229.8 F 0.0
20:00 232.3 F 0.0
25:00 226.1 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 19:05
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

04/19/2023 11:31 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____