

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1773

Name: RITTER 1  
Sterilizer ID: V2178198  
03/31/2022 07:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.3 F 0.0  
2:00 149.4 F 0.3  
4:00 163.2 F 2.2  
6:00 175.9 F 4.9  
8:00 195.5 F 8.5  
10:00 211.8 F 12.8  
12:00 222.0 F 6.5  
14:00 232.5 F 11.0  
16:00 230.4 F 7.3  
18:00 244.0 F 13.2  
20:00 241.7 F 11.4  
22:00 254.1 F 17.7  
24:00 265.2 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.1 F 28.9  
2:00 272.1 F 29.2  
3:00 271.7 F 28.6  
4:00 272.5 F 29.3

Min 271.4 F 28.4  
Max 272.5 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 205.4 F 0.0  
10:00 219.0 F 0.0  
15:00 240.0 F 0.0  
20:00 228.9 F 0.0  
25:00 229.8 F 0.0  
30:00 227.2 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 25:50  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:03:51  
CYCLE COMPLETE

03/31/2022 08:15 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_