
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001659
11:45 11/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:36
PRESSURIZING 1:41
STERILIZING 2:42
136.4 C 220kPa 2:42
136.4 C 226kPa 3:12
136.4 C 225kPa 3:42
136.3 C 225kPa 4:12
136.4 C 226kPa 4:42
136.3 C 225kPa 5:12
136.2 C 222kPa 5:42
136.2 C 224kPa 6:12
136.2 C 224kPa 6:12
VENTING 6:12
AIR DRYING 7:02
CYCLE COMPLETE 10:16

Digital Signature #
0E7D3F2847EE6215
