
Midmark M11 - v1.0.5
Total Cycles: 2078

Name: RITTER2
Sterilizer ID: V2313228
02/22/2023 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.4 F 0.0
2:00 183.2 F 3.4
4:00 201.8 F 8.5
6:00 220.3 F 9.0
8:00 231.7 F 10.8
10:00 234.6 F 9.0
12:00 238.1 F 9.5
14:00 250.4 F 15.8
16:00 263.3 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.2 F 29.3
2:00 271.6 F 29.3
3:00 272.1 F 29.2
4:00 271.5 F 29.2

Min 271.2 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 231.7 F 0.0
10:00 225.5 F 0.0
15:00 223.4 F 0.0
20:00 220.9 F 0.0
25:00 220.2 F 0.0
30:00 219.7 F 0.0

Drying Complete

FILLING: 2:01
HEATING: 17:60
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:14
CYCLE COMPLETE

02/22/2023 03:07 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____