

-----  
Midmark M11 - v1.0.5  
Total Cycles: 296

Name: RITTER2  
Sterilizer ID: V2313228  
08/20/2021 11:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.1 F 0.0  
2:00 144.2 F 2.0  
4:00 163.3 F 4.8  
6:00 186.9 F 8.5  
8:00 202.8 F 12.6  
10:00 214.7 F 6.6  
12:00 226.8 F 11.2  
14:00 229.8 F 7.9  
16:00 241.5 F 13.6  
18:00 242.1 F 12.4  
20:00 253.7 F 18.5  
22:00 264.3 F 24.9  
24:00 271.8 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.8  
1:00 272.5 F 30.6  
2:00 272.1 F 30.1  
3:00 271.3 F 29.7  
4:00 271.7 F 30.4

Min 271.2 F 29.6  
Max 272.8 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.8 F 0.0  
5:00 204.6 F 0.0  
10:00 201.7 F 0.0  
15:00 195.1 F 0.0  
20:00 188.9 F 0.0  
25:00 188.0 F 0.0  
30:00 187.5 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 24:07  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:02:52  
CYCLE COMPLETE

08/20/2021 12:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_