
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002431
10:44 13/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:35
STERILIZING 3:37
136.5 C 224kPa 3:37
136.4 C 225kPa 4:07
136.4 C 221kPa 4:37
136.3 C 223kPa 5:07
136.3 C 224kPa 5:37
136.4 C 221kPa 6:07
136.4 C 223kPa 6:37
136.4 C 223kPa 7:07
136.4 C 223kPa 7:07
VENTING 7:07
AIR DRYING 8:00
CYCLE COMPLETE 8:02

Digital Signature #
9D4240384CB0554B
