
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001456
7:14 06/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:02
STERILIZING 4:07
136.4 C 220kPa 4:07
136.4 C 223kPa 4:37
136.0 C 220kPa 5:07
136.3 C 223kPa 5:37
136.4 C 223kPa 6:07
136.4 C 223kPa 6:37
136.4 C 223kPa 7:07
136.4 C 223kPa 7:37
136.4 C 223kPa 7:37
VENTING 7:37
AIR DRYING 8:29
CYCLE COMPLETE 8:36

Digital Signature #
A6581396EDEB4427
