

-----  
Midmark M11 - v1.0.5  
Total Cycles: 264

Name: RITTER2  
Sterilizer ID: V2313228  
08/05/2021 06:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.0 F 0.0  
2:00 73.2 F 0.0  
4:00 79.5 F 0.1  
6:00 107.0 F 1.4  
8:00 137.0 F 4.2  
10:00 161.0 F 7.6  
12:00 185.8 F 11.6  
14:00 211.9 F 4.3  
16:00 214.5 F 8.0  
18:00 226.8 F 11.9  
20:00 231.8 F 7.8  
22:00 242.4 F 12.9  
24:00 240.3 F 10.9  
26:00 252.4 F 16.9  
28:00 262.6 F 23.0  
30:00 272.6 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 271.4 F 28.9  
2:00 271.3 F 28.8  
3:00 271.4 F 29.1  
4:00 272.4 F 29.3

Min 271.0 F 28.8  
Max 272.4 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.9 F 0.0  
5:00 223.8 F 0.0  
10:00 214.9 F 0.0  
15:00 201.8 F 0.0  
20:00 193.4 F 0.0  
25:00 196.7 F 0.0  
30:00 184.7 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 30:17  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:08:20  
CYCLE COMPLETE

08/05/2021 07:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_