

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2077

Name: RITTER 1  
Sterilizer ID: V2178198  
06/29/2022 04:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
2:00 183.1 F 3.7  
4:00 189.1 F 8.1  
6:00 222.7 F 13.2  
8:00 229.0 F 8.3  
10:00 241.2 F 12.6  
12:00 239.1 F 10.6  
14:00 235.2 F 8.3  
16:00 250.2 F 15.2  
18:00 261.9 F 21.8  
20:00 272.6 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.7  
1:00 272.1 F 28.7  
2:00 272.7 F 29.0  
3:00 272.2 F 29.2  
4:00 271.7 F 28.6

Min 271.6 F 28.6  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.0 F 0.0  
5:00 217.0 F 0.0  
10:00 235.4 F 0.0  
15:00 233.7 F 0.0  
20:00 226.3 F 0.0  
25:00 223.5 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 21:08  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:33  
CYCLE COMPLETE

06/29/2022 05:03 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_