
Midmark M11 - v1.0.5
Total Cycles: 1299

Name: RITTER2
Sterilizer ID: V2313228
07/14/2022 12:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.9 F 0.0
2:00 133.4 F 0.0
4:00 158.4 F 0.6
6:00 189.6 F 3.4
8:00 219.8 F 7.8
10:00 222.0 F 12.7
12:00 229.5 F 8.2
14:00 240.3 F 13.2
16:00 240.1 F 11.1
18:00 237.7 F 9.7
20:00 251.5 F 16.4
22:00 263.0 F 23.2
24:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.9 F 29.4
2:00 272.7 F 29.8
3:00 271.7 F 29.3
4:00 273.0 F 29.5

Min 271.6 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.2 F 0.0
5:00 235.3 F 0.0
10:00 215.5 F 0.0
15:00 211.5 F 0.0
20:00 198.0 F 0.0
25:00 203.2 F 0.0
30:00 160.3 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 24:55
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:03:17
CYCLE COMPLETE

07/14/2022 01:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____