
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002077
8:53 23/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:09
STERILIZING 4:16
136.4 C 225kPa 4:16
135.9 C 222kPa 4:46
136.3 C 224kPa 5:16
136.5 C 225kPa 5:46
136.4 C 224kPa 6:16
136.4 C 225kPa 6:46
136.4 C 223kPa 7:16
136.2 C 221kPa 7:46
136.2 C 221kPa 7:46
VENTING 7:46
AIR DRYING 8:41
CYCLE COMPLETE 9:02

Digital Signature #
362A967897A2A555
