
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002533
14:57 21/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:53
STERILIZING 3:59
136.0 C 221kPa 3:59
136.2 C 223kPa 4:29
136.4 C 225kPa 4:59
136.3 C 224kPa 5:29
136.4 C 225kPa 5:59
136.2 C 224kPa 6:29
136.4 C 225kPa 6:59
136.4 C 224kPa 7:29
136.4 C 224kPa 7:29
VENTING 7:29
AIR DRYING 8:21
CYCLE COMPLETE 38:22

Digital Signature #
B1C9703BC2E00AF3
