
Midmark M11 - v1.0.5
Total Cycles: 4050

Name: RITTER2
Sterilizer ID: V2313228
10/23/2024 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.6 F 0.0
2:00 162.0 F 2.3
4:00 169.9 F 6.3
6:00 193.8 F 11.0
8:00 213.9 F 5.0
10:00 223.5 F 10.3
12:00 228.4 F 7.6
14:00 240.5 F 12.8
16:00 240.2 F 11.3
18:00 253.1 F 17.8
20:00 264.5 F 24.5
22:00 271.2 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.4 F 30.0
2:00 271.8 F 29.3
3:00 271.7 F 29.8
4:00 272.4 F 29.8

Min 271.2 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 198.7 F 0.0
10:00 196.8 F 0.0
15:00 194.3 F 0.0
20:00 195.8 F 0.0
25:00 220.4 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

10/23/2024 05:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____