
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002346
12:27 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:09
STERILIZING 4:21
136.0 C 223kPa 4:21
136.4 C 224kPa 4:51
136.2 C 224kPa 5:21
136.3 C 222kPa 5:51
136.0 C 223kPa 6:21
136.3 C 225kPa 6:51
136.4 C 225kPa 7:21
136.3 C 225kPa 7:51
136.3 C 225kPa 7:51
VENTING 7:51
AIR DRYING 8:45
CYCLE COMPLETE 38:45

Digital Signature #
51F0C3332771706A
