

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3753

Name: RITTER 1  
Sterilizer ID: V2178198  
09/11/2023 04:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.3 F 0.0  
2:00 166.5 F 0.0  
4:00 180.5 F 3.0  
6:00 200.3 F 6.2  
8:00 216.0 F 10.5  
10:00 223.0 F 6.6  
12:00 232.3 F 9.6  
14:00 237.2 F 10.0  
16:00 242.8 F 12.3  
18:00 241.6 F 11.0  
20:00 253.5 F 17.1  
22:00 264.0 F 23.3  
24:00 272.1 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.6 F 29.0  
2:00 271.8 F 28.3  
3:00 272.4 F 29.1  
4:00 272.2 F 28.6

Min 271.5 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.8 F 0.0  
5:00 199.0 F 0.0  
10:00 208.2 F 0.0  
15:00 206.3 F 0.0  
20:00 204.2 F 0.0  
25:00 203.3 F 0.0  
30:00 202.7 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 24:02  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:03:03  
CYCLE COMPLETE

09/11/2023 05:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_