
Midmark M11 - v1.0.5
Total Cycles: 2657

Name: RITTER2
Sterilizer ID: V2313228
07/17/2023 04:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.5 F 0.0
2:00 155.4 F 0.8
4:00 173.9 F 4.2
6:00 198.7 F 8.4
8:00 214.1 F 12.7
10:00 223.5 F 7.5
12:00 233.6 F 12.4
14:00 235.9 F 9.9
16:00 237.0 F 9.2
18:00 248.5 F 15.0
20:00 260.0 F 21.5
22:00 270.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.1 F 29.6
2:00 271.5 F 29.5
3:00 272.2 F 29.6
4:00 271.2 F 29.4

Min 271.2 F 29.1
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 219.7 F 0.0
10:00 201.4 F 0.0
15:00 198.8 F 0.0
20:00 197.0 F 0.0
25:00 194.2 F 0.0
30:00 194.0 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 22:49
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:01:37
CYCLE COMPLETE

07/17/2023 05:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____