
Midmark M11 - v1.0.0
Total Cycles: 4197

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2024 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 71.1 F 0.3
4:00 79.5 F 1.2
6:00 106.4 F 2.9
8:00 136.3 F 5.1
10:00 159.8 F 7.7
12:00 179.0 F 10.7
14:00 200.3 F 8.8
16:00 205.3 F 4.2
18:00 214.3 F 7.2
20:00 224.2 F 10.6
22:00 229.8 F 8.2
24:00 236.4 F 10.2
26:00 236.7 F 9.2
28:00 245.0 F 12.9
30:00 255.8 F 18.4
32:00 265.2 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.1 F 29.0
2:00 272.0 F 28.7
3:00 271.2 F 28.5
4:00 272.4 F 29.1

Min 271.1 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.8 F 0.0
5:00 228.6 F 0.0
10:00 233.7 F 0.0
15:00 227.3 F 0.0
20:00 227.8 F 0.0
25:00 229.4 F 0.0
30:00 230.4 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 33:56
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:12:30
CYCLE COMPLETE

01/12/2024 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____