
Midmark M11 - v1.0.5
Total Cycles: 777

Name: RITTER2
Sterilizer ID: V2313228
02/10/2022 07:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 74.5 F 0.0
4:00 86.7 F 0.0
6:00 128.1 F 0.1
8:00 160.3 F 2.8
10:00 182.0 F 6.1
12:00 200.5 F 9.8
14:00 212.7 F 10.8
16:00 223.2 F 7.5
18:00 232.7 F 11.9
20:00 233.9 F 8.6
22:00 241.1 F 11.5
24:00 244.2 F 12.9
26:00 256.0 F 19.2
28:00 266.6 F 25.6
30:00 272.7 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.9 F 29.9
2:00 271.8 F 29.7
3:00 273.0 F 29.8
4:00 271.8 F 29.7

Min 271.5 F 29.3
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.1 F 0.0
5:00 240.4 F 0.0
10:00 200.5 F 0.0
15:00 181.2 F 0.0
20:00 168.6 F 0.0
25:00 188.9 F 0.0
30:00 198.0 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 30:25
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

02/10/2022 08:59 AM
ACCEPTED: X
REJECTED:
Operator: U02